

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044128.D
 Acq On : 08 Jun 2021 21:15
 Operator : SY/MD
 Sample : M2612-39
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(78-82)

Quant Time: Jun 08 23:15:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	212919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	429643	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	460127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	211813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	193629	52.217	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.440%	
35) Dibromofluoromethane	5.298	113	152723	49.812	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.620%	
50) Toluene-d8	7.906	98	597773	50.546	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.100%	
62) 4-Bromofluorobenzene	10.639	95	220603	47.115	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.220%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	16527	7.722	ug/l	100
20) Methylene Chloride	3.054	84	12554	2.290	ug/l	97
43) Isopropyl Acetate	5.919	43	25189	2.921	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044128.D
Acq On : 08 Jun 2021 21:15
Operator : SY/MD
Sample : M2612-39
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(78-82)

Quant Time: Jun 08 23:15:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

