



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044113.D
 Acq On : 08 Jun 2021 15:15
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Title : SW846 8260
 Last Update : Tue Jun 08 17:38:51 2021



AutoFind: Scans 2892, 2893, 2894; Background Corrected with Scan 2879

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	7315	PASS
75	95	30	60	52.0	20603	PASS
95	95	100	100	100.0	39616	PASS
96	95	5	9	7.5	2982	PASS
173	174	0.00	2	1.4	411	PASS
174	95	50	100	75.5	29896	PASS
175	174	5	9	6.6	1971	PASS
176	174	95	101	97.8	29245	PASS
177	176	5	9	6.7	1961	PASS