

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
 Data File : VU048956.D
 Acq On : 08 Jun 2022 12:10
 Operator : SY/MD
 Sample : N3194-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-1

Quant Time: Jun 08 12:56:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	241071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	405154	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	384929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	175710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	180423	49.835	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.660%		
35) Dibromofluoromethane	5.289	113	136112	47.090	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.180%		
50) Toluene-d8	7.896	98	503390	48.584	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.160%		
62) 4-Bromofluorobenzene	10.632	95	196932	50.257	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.520%		
Target Compounds					Qvalue	
16) Acetone	2.623	43	46361	25.127	ug/l	98
20) Methylene Chloride	3.041	84	12413	3.956	ug/l	98
25) 2-Butanone	4.716	43	7457	2.813	ug/l	99
43) Isopropyl Acetate	5.912	43	26180	3.382	ug/l	99
84) 1,2,4-Trimethylbenzene	11.468	105	1608	1.335	ug/l	91
89) n-Butylbenzene	12.211	91	4755	2.632	ug/l	98
95) Naphthalene	14.092	128	18664	3.814	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
Data File : VU048956.D
Acq On : 08 Jun 2022 12:10
Operator : SY/MD
Sample : N3194-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-1

Quant Time: Jun 08 12:56:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

