

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
 Data File : VU048957.D
 Acq On : 08 Jun 2022 12:34
 Operator : SY/MD
 Sample : N3194-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-3

Quant Time: Jun 08 13:11:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	248594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	410732	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	386058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	163688	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	183530	49.159	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.320%
35) Dibromofluoromethane	5.289	113	140673	48.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.020%
50) Toluene-d8	7.896	98	519258	49.435	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	10.632	95	176139	44.340	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.680%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	48162	25.313	ug/l	100
18) Methyl Acetate	2.948	43	7128	1.492	ug/l	95
20) Methylene Chloride	3.041	84	13200	4.080	ug/l	91
25) 2-Butanone	4.710	43	7573	2.771	ug/l	90
43) Isopropyl Acetate	5.912	43	27914	3.557	ug/l	97
89) n-Butylbenzene	12.211	91	1043	2.291	ug/l	92
95) Naphthalene	14.095	128	7481	3.084	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
Data File : VU048957.D
Acq On : 08 Jun 2022 12:34
Operator : SY/MD
Sample : N3194-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-3

Quant Time: Jun 08 13:11:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

