

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
 Data File : VU048961.D
 Acq On : 08 Jun 2022 14:43
 Operator : SY/MD
 Sample : N3143-05RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20220425-SPLP-AMENDED-COMP-ERE

Quant Time: Jun 08 15:34:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	239483	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	398759	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	362669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	162640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	185299	51.521	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.040%
35) Dibromofluoromethane	5.289	113	71170	25.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	50.040%#
50) Toluene-d8	7.896	98	490275	48.077	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.160%
62) 4-Bromofluorobenzene	10.632	95	175386	45.476	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.960%
Target Compounds						
						Qvalue
16) Acetone	2.623	43	64615	35.253	ug/l	99
20) Methylene Chloride	3.044	84	4595	1.474	ug/l	92
25) 2-Butanone	4.710	43	14396	5.467	ug/l	98
84) 1,2,4-Trimethylbenzene	11.472	105	795	1.266	ug/l #	62
95) Naphthalene	14.092	128	7585	3.095	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
Data File : VU048961.D
Acq On : 08 Jun 2022 14:43
Operator : SY/MD
Sample : N3143-05RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20220425-SPLP-AMENDED-COMP-ERE

Quant Time: Jun 08 15:34:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

