

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
 Data File : VU048962.D
 Acq On : 08 Jun 2022 15:06
 Operator : SY/MD
 Sample : N3143-06RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20220427-SPLP-AMENDED-SB-8-(DISCRETE-COMP-G)RE

Quant Time: Jun 08 15:52:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	248883	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	417160	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	376151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	167197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	188667	50.476	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.960%
35) Dibromofluoromethane	5.288	113	55649	18.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	37.400%#
50) Toluene-d8	7.896	98	498405	46.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.440%
62) 4-Bromofluorobenzene	10.632	95	176966	43.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.720%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	44907	23.575	ug/l	99
20) Methylene Chloride	3.044	84	5128	1.583	ug/l	95
25) 2-Butanone	4.710	43	11244	4.109	ug/l	99
27) cis-1,2-Dichloroethene	4.668	96	6095	1.817	ug/l	90
95) Naphthalene	14.098	128	3936	2.817	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
Data File : VU048962.D
Acq On : 08 Jun 2022 15:06
Operator : SY/MD
Sample : N3143-06RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20220427-SPLP-AMENDED-SB-8-(DISCRETE-COMP-G)RE

Quant Time: Jun 08 15:52:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

