

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
 Data File : VU048966.D
 Acq On : 08 Jun 2022 16:41
 Operator : SY/MD
 Sample : N3209-08
 Misc : 5.03g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-14B-10

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/09/2022
 Supervised By :Mahesh Dadoda 06/09/2022

Quant Time: Jun 09 09:56:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	261937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	439181	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	411367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	202633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	197713	50.260	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.520%			
35) Dibromofluoromethane	5.289	113	147559	47.095	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.200%			
50) Toluene-d8	7.896	98	540707	48.142	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 96.280%			
62) 4-Bromofluorobenzene	10.639	95	211140	49.708	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.420%			
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.173	59	19272	19.324	ug/l #	30
25) 2-Butanone	4.726	43	31736	11.019	ug/l	98
39) Methylcyclohexane	6.729	83	15597m	3.253	ug/l	
73) Isopropylbenzene	10.488	105	37001	2.828	ug/l	97
78) n-propylbenzene	10.909	91	48642	3.120	ug/l	97
85) sec-Butylbenzene	11.648	105	41313	3.024	ug/l	99
89) n-Butylbenzene	12.211	91	22113	3.986	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
Data File : VU048966.D
Acq On : 08 Jun 2022 16:41
Operator : SY/MD
Sample : N3209-08
Misc : 5.03g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-14B-10

Quant Time: Jun 09 09:56:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/09/2022
Supervised By :Mahesh Dadoda 06/09/2022

