

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
 Data File : VU048970.D
 Acq On : 08 Jun 2022 18:16
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Jun 09 09:57:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	398511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	685935	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	719426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	418665	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	275949	46.108	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.220%
35) Dibromofluoromethane	5.289	113	228713	46.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.480%
50) Toluene-d8	7.896	98	849159	48.408	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	10.632	95	376366	56.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060822\
Data File : VU048970.D
Acq On : 08 Jun 2022 18:16
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Jun 09 09:57:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

