

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060920\
 Data File : VU038694.D
 Acq On : 09 Jun 2020 01:51
 Operator : SY/MD
 Sample : L2896-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW0K4

Quant Time: Jun 09 07:29:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 09 06:12:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	557060	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	537771	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	249721	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	172467	55.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.82%
7) Chloroethane-d5	1.93	69	138150	57.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	114.44%
11) 1,1-Dichloroethene-d2	2.59	63	268707	39.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.22%
21) 2-Butanone-d5	4.66	46	320248	105.77	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.77%
24) Chloroform-d	5.10	84	339323	50.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.14%
26) 1,2-Dichloroethane-d4	5.74	65	230886	51.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.82%
32) Benzene-d6	5.76	84	694059	53.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.16%
36) 1,2-Dichloropropane-d6	6.72	67	210797	52.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.60%
41) Toluene-d8	7.92	98	600754	51.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.64%
43) trans-1,3-Dichloropropene-	8.20	79	97319	48.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.84%
47) 2-Hexanone-d5	8.65	63	232890	109.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.59%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	300814	50.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.02%
64) 1,2-Dichlorobenzene-d4	12.20	152	235126	54.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.46%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060920\
Data File : VU038694.D
Acq On : 09 Jun 2020 01:51
Operator : SY/MD
Sample : L2896-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW0K4

Quant Time: Jun 09 07:29:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 09 06:12:02 2020
Response via : Initial Calibration

