

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060922\
 Data File : VU048974.D
 Acq On : 09 Jun 2022 12:56
 Operator : SY/MD
 Sample : VU0609WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK139

Quant Time: Jun 10 01:02:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 10 01:01:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	334537	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	327989	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	158558	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	107245	52.340	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 104.680%		
7) Chloroethane-d5	1.893	69	78273	50.989	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	= 101.980%		
11) 1,1-Dichloroethene-d2	2.559	63	145896	31.484	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 62.960%		
21) 2-Butanone-d5	4.620	46	203824	104.531	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 104.530%		
24) Chloroform-d	5.060	84	219444	42.381	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 84.760%		
26) 1,2-Dichloroethane-d4	5.700	65	143969	39.412	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 78.820%		
32) Benzene-d6	5.726	84	423690	46.416	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 92.840%		
36) 1,2-Dichloropropane-d6	6.690	67	141482	52.796	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 105.600%		
41) Toluene-d8	7.899	98	382698	44.572	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 89.140%		
43) trans-1,3-Dichloroprop...	8.179	79	70564	41.179	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 82.360%		
47) 2-Hexanone-d5	8.632	63	145924	99.294	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 99.290%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	228459	51.521	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 103.040%		
66) 1,2-Dichlorobenzene-d4	12.195	152	158861	49.010	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 98.020%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060922\
Data File : VU048974.D
Acq On : 09 Jun 2022 12:56
Operator : SY/MD
Sample : VU0609WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK139

Quant Time: Jun 10 01:02:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 10 01:01:44 2022
Response via : Initial Calibration

