

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060922\
 Data File : VU048978.D
 Acq On : 09 Jun 2022 14:43
 Operator : SY/MD
 Sample : VIBLK073
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK073

Quant Time: Jun 10 01:03:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 10 01:01:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	305549	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	289092	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	135534	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	94729	50.618	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 101.240%		
7) Chloroethane-d5	1.893	69	72948	52.028	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	= 104.060%		
11) 1,1-Dichloroethene-d2	2.562	63	129108	30.504	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 61.000%		
21) 2-Butanone-d5	4.620	46	227128	127.533	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 127.530%		
24) Chloroform-d	5.060	84	212826	45.003	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 90.000%		
26) 1,2-Dichloroethane-d4	5.700	65	143450	42.996	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 86.000%		
32) Benzene-d6	5.726	84	395187	49.118	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 98.240%		
36) 1,2-Dichloropropane-d6	6.690	67	136985	57.996	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 116.000%		
41) Toluene-d8	7.896	98	354305	46.817	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 93.640%		
43) trans-1,3-Dichloroprop...	8.179	79	65858	43.604	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 87.200%		
47) 2-Hexanone-d5	8.632	63	158116	122.066	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 122.070%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	218034	55.786	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 111.580%		
66) 1,2-Dichlorobenzene-d4	12.192	152	142242	51.338	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 102.680%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060922\
Data File : VU048978.D
Acq On : 09 Jun 2022 14:43
Operator : SY/MD
Sample : VIBLK073
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK073

Quant Time: Jun 10 01:03:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 10 01:01:44 2022
Response via : Initial Calibration

