

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060923\
 Data File : VU054460.D
 Acq On : 09 Jun 2023 14:10
 Operator : MD/SY
 Sample : 03105-19
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FH0

Quant Time: Jun 12 01:00:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 12 00:58:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	293397	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	278566	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	134261	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	64530	3.456	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.200%
7) Chloroethane-d5	1.914	69	69766	3.991	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	2.567	65	30592	3.668	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.400%
20) 2-Butanone-d5	4.631	46	266680	50.798	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.600%
24) Chloroform-d	5.059	84	170759	4.375	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
26) 1,2-Dichloroethane-d4	5.699	65	88021	4.626	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.724	84	325672	4.206	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.686	67	117575	4.547	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.000%
41) Toluene-d8	7.895	98	275862	4.014	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	8.178	79	38173	4.245	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.631	63	169823	45.012	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.020%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	90136	4.674	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	109464	4.834	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060923\
Data File : VU054460.D
Acq On : 09 Jun 2023 14:10
Operator : MD/SY
Sample : 03105-19
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0FH0

Quant Time: Jun 12 01:00:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 12 00:58:03 2023
Response via : Initial Calibration

