

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU060923\  
 Data File : VU054469.D  
 Acq On : 09 Jun 2023 17:48  
 Operator : MD/SY  
 Sample : 03118-07  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 C0FL5

Quant Time: Jun 12 01:02:24 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR060723WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Jun 12 00:58:03 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 6.245  | 114   | 325902   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 9.415  | 117   | 310893   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.807 | 152   | 146076   | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.599  | 65    | 60989    | 2.941    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 58.800%  |
| 7) Chloroethane-d5            | 1.914  | 69    | 68211    | 3.513    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 70.200%  |
| 11) 1,1-Dichloroethene-d2     | 2.563  | 65    | 29667    | 3.203    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 64.000%  |
| 20) 2-Butanone-d5             | 4.631  | 46    | 272967   | 46.810   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 93.620%  |
| 24) Chloroform-d              | 5.058  | 84    | 172325   | 3.975    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 79.400%  |
| 26) 1,2-Dichloroethane-d4     | 5.698  | 65    | 89719    | 4.245    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 85.000%  |
| 32) Benzene-d6                | 5.724  | 84    | 324889   | 3.759    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 75.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.688  | 67    | 119270   | 4.133    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 82.600%  |
| 41) Toluene-d8                | 7.894  | 98    | 272382   | 3.551    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 71.000%  |
| 43) trans-1,3-Dichloroprop... | 8.177  | 79    | 37157    | 3.702    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 74.000%  |
| 46) 2-Hexanone-d5             | 8.630  | 63    | 164973   | 39.180   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 78.360%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.753 | 84    | 93378    | 4.339    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 86.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 12.190 | 152   | 109709   | 4.453    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 89.000%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU060923\  
Data File : VU054469.D  
Acq On : 09 Jun 2023 17:48  
Operator : MD/SY  
Sample : 03118-07  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
C0FL5

Quant Time: Jun 12 01:02:24 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR060723WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Mon Jun 12 00:58:03 2023  
Response via : Initial Calibration

