

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061019\
 Data File : VU032554.D
 Acq On : 10 Jun 2019 09:26
 Operator : JC/SP
 Sample : VU0610WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK54

Quant Time: Jun 11 04:55:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 11 04:40:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	208962	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	193904	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	100623	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	75951	56.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	112.04%
7) Chloroethane-d5	1.68	69	59746	54.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.68%
11) 1,1-Dichloroethene-d2	2.27	63	110553	41.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.30%
21) 2-Butanone-d5	4.17	46	116830	102.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.97%
24) Chloroform-d	4.64	84	133594	52.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.68%
26) 1,2-Dichloroethane-d4	5.30	65	89198	56.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.00%
32) Benzene-d6	5.33	84	280112	55.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.30%
36) 1,2-Dichloropropane-d6	6.32	67	88367	55.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.96%
41) Toluene-d8	7.56	98	262059	54.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.80%
43) trans-1,3-Dichloropropene-	7.85	79	42847	56.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	113.44%
47) 2-Hexanone-d5	8.31	63	78179	102.11	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.11%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	118942	52.00	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.00%
64) 1,2-Dichlorobenzene-d4	11.85	152	104403	54.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061019\
Data File : VU032554.D
Acq On : 10 Jun 2019 09:26
Operator : JC/SP
Sample : VU0610WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK54

Quant Time: Jun 11 04:55:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 11 04:40:52 2019
Response via : Initial Calibration

