

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061019\
 Data File : VU032556.D
 Acq On : 10 Jun 2019 11:04
 Operator : JC/SP
 Sample : K3277-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB8B0

Quant Time: Jun 11 05:10:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 11 04:40:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	214225	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	199945	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	104302	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	67524	48.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.16%
7) Chloroethane-d5	1.68	69	53412	47.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.64%
11) 1,1-Dichloroethene-d2	2.27	63	100408	36.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.80%
21) 2-Butanone-d5	4.17	46	127112	109.29	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.29%
24) Chloroform-d	4.64	84	121306	46.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.30	65	79971	48.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.94%
32) Benzene-d6	5.33	84	255234	49.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.34%
36) 1,2-Dichloropropane-d6	6.32	67	79810	49.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.06%
41) Toluene-d8	7.56	98	236267	48.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.00%
43) trans-1,3-Dichloropropene-	7.85	79	39344	50.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.02%
47) 2-Hexanone-d5	8.31	63	82220	104.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.14%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	114923	48.73	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.46%
64) 1,2-Dichlorobenzene-d4	11.85	152	98772	49.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.88%

Target Compounds					Ovalue
13) Acetone	2.32	43	12016	10.012 ug/L	99
34) Trichloroethene	6.18	95	118858	77.766 ug/L	99
42) Toluene	7.64	91	8481	1.378 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061019\
Data File : VU032556.D
Acq On : 10 Jun 2019 11:04
Operator : JC/SP
Sample : K3277-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB8B0

Quant Time: Jun 11 05:10:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 11 04:40:52 2019
Response via : Initial Calibration

