

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061019\
 Data File : VU032569.D
 Acq On : 10 Jun 2019 16:28
 Operator : JC/SP
 Sample : VU0610WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK55

Quant Time: Jun 11 05:57:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 11 05:38:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	199722	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	186684	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	97317	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	63161	48.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.48%
7) Chloroethane-d5	1.68	69	51399	49.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.72%
11) 1,1-Dichloroethene-d2	2.27	63	93235	36.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.50%
21) 2-Butanone-d5	4.17	46	107837	99.45	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.45%
24) Chloroform-d	4.64	84	114873	47.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.06%
26) 1,2-Dichloroethane-d4	5.30	65	76217	50.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.12%
32) Benzene-d6	5.33	84	238875	49.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.58%
36) 1,2-Dichloropropane-d6	6.32	67	75677	49.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.60%
41) Toluene-d8	7.56	98	222687	48.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.92%
43) trans-1,3-Dichloropropene-	7.85	79	36984	50.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.70%
47) 2-Hexanone-d5	8.31	63	70925	96.22	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.22%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	105353	47.84	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.68%
64) 1,2-Dichlorobenzene-d4	11.85	152	91857	49.28	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061019\
Data File : VU032569.D
Acq On : 10 Jun 2019 16:28
Operator : JC/SP
Sample : VU0610WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK55

Quant Time: Jun 11 05:57:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 11 05:38:55 2019
Response via : Initial Calibration

