

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061020\
 Data File : VU038759.D
 Acq On : 10 Jun 2020 05:49
 Operator : SY/MD
 Sample : L2913-08 100X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D95

Quant Time: Jun 10 07:43:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 10 06:36:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	521818	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	493869	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	225085	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	129927	44.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.14%
7) Chloroethane-d5	1.93	69	116970	51.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.44%
11) 1,1-Dichloroethene-d2	2.59	63	216817	34.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.24%
21) 2-Butanone-d5	4.66	46	330196	116.42	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.42%
24) Chloroform-d	5.10	84	323936	51.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.06%
26) 1,2-Dichloroethane-d4	5.74	65	227787	54.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.36%
32) Benzene-d6	5.76	84	624779	52.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.04%
36) 1,2-Dichloropropane-d6	6.72	67	200606	54.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.44%
41) Toluene-d8	7.92	98	524443	48.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.56%
43) trans-1,3-Dichloropropene-	8.20	79	92837	50.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.58%
47) 2-Hexanone-d5	8.65	63	234331	120.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.07%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	298266	54.54	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.08%
64) 1,2-Dichlorobenzene-d4	12.20	152	217824	55.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.48%

Target Compounds					Ovalue
33) Benzene	5.81	78	755789	51.534 ug/L	100
52) Ethylbenzene	9.59	91	28320	1.673 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061020\
Data File : VU038759.D
Acq On : 10 Jun 2020 05:49
Operator : SY/MD
Sample : L2913-08 100X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D95

Quant Time: Jun 10 07:43:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 10 06:36:11 2020
Response via : Initial Calibration

