

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061020\
 Data File : VU038765.D
 Acq On : 10 Jun 2020 09:06
 Operator : SY/MD
 Sample : VIBLK43
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK43

Quant Time: Jun 11 01:05:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 00:28:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	387967	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	377200	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	180452	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	67284	31.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.08%
7) Chloroethane-d5	1.93	69	56138	33.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	66.78%#
11) 1,1-Dichloroethene-d2	2.59	63	109731	23.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	46.46%#
21) 2-Butanone-d5	4.66	46	136350	64.66	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	64.66%
24) Chloroform-d	5.10	84	144984	31.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	62.04%#
26) 1,2-Dichloroethane-d4	5.74	65	102369	33.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	66.10%#
32) Benzene-d6	5.76	84	298603	32.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	65.72%#
36) 1,2-Dichloropropane-d6	6.72	67	91156	32.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	65.10%#
41) Toluene-d8	7.93	98	250919	30.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	61.12%#
43) trans-1,3-Dichloropropene-	8.20	79	42167	29.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.82%#
47) 2-Hexanone-d5	8.65	63	91756	61.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	61.56%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	127482	30.52	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	61.04%#
64) 1,2-Dichlorobenzene-d4	12.21	152	105894	33.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.60%#


Target Compounds					Ovalue
46) Tetrachloroethene	8.57	164	2494	1.213 ug/L	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061020\
Data File : VU038765.D
Acq On : 10 Jun 2020 09:06
Operator : SY/MD
Sample : VIBLK43
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK43

Quant Time: Jun 11 01:05:10 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 00:28:57 2020
Response via : Initial Calibration





#46

Tetrachloroethene

Concen: 1.213 ug/L

RT: 8.57 min Scan# 2250

Delta R.T. 0.00 min

Lab File: VU038765.D


Acq: 10 Jun 2020 09:06

Instrument :

MSVOA_U

ClientSampleId :

VIBLK43



Tot Ion:164 Resp: 2494

Ion Ratio Lower Upper

164 100

129 91.3 67.2 124.8

131 84.8 65.5 121.6

166 100.3 90.4 168.0

Abundance Ion 164.00 (163.70 to 164.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 131.00 (130.70 to 131.70): V

Ion 166.00 (165.70 to 166.70): V

