

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061021\
 Data File : VU044142.D
 Acq On : 10 Jun 2021 09:46
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 01:17:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	47.937	4.1	102	0.00
3 P	Chloromethane	50.000	51.005	-2.0	102	0.00
4 C	Vinyl Chloride	50.000	46.200	7.6#	101	0.00
5 T	Bromomethane	50.000	55.693	-11.4	104	0.00
6 T	Chloroethane	50.000	49.437	1.1	100	0.00
7 T	Trichlorofluoromethane	50.000	46.617	6.8	101	0.00
8 T	Diethyl Ether	50.000	47.538	4.9	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.569	6.9	102	0.00
10 T	Methyl Iodide	50.000	49.377	1.2	96	0.00
11 T	Tert butyl alcohol	250.000	235.627	5.7	104	0.00
12 CM	1,1-Dichloroethene	50.000	46.939	6.1#	102	0.00
13 T	Acrolein	250.000	218.201	12.7	89	0.00
14 T	Allyl chloride	50.000	47.980	4.0	103	0.00
15 T	Acrylonitrile	250.000	240.514	3.8	99	0.00
16 T	Acetone	250.000	295.706	-18.3	123	0.00
17 T	Carbon Disulfide	50.000	46.119	7.8	102	0.00
18 T	Methyl Acetate	50.000	48.666	2.7	102	0.00
19 T	Methyl tert-butyl Ether	50.000	49.967	0.1	103	0.00
20 T	Methylene Chloride	50.000	50.299	-0.6	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.308	5.4	101	0.00
22 T	Diisopropyl ether	50.000	50.540	-1.1	104	0.00
23 T	Vinyl Acetate	250.000	255.097	-2.0	103	0.00
24 P	1,1-Dichloroethane	50.000	47.153	5.7	102	0.00
25 T	2-Butanone	250.000	248.929	0.4	105	0.00
26 T	2,2-Dichloropropane	50.000	47.791	4.4	104	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.654	4.7	103	0.00
28 T	Bromochloromethane	50.000	49.686	0.6	107	0.00
29 T	Tetrahydrofuran	250.000	243.679	2.5	99	0.00
30 C	Chloroform	50.000	47.077	5.8#	102	0.00
31 T	Cyclohexane	50.000	46.970	6.1	103	0.00
32 T	1,1,1-Trichloroethane	50.000	47.152	5.7	102	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.300	5.4	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	46.397	7.2	98	0.00
36 T	1,1-Dichloropropene	50.000	47.596	4.8	103	0.00
37 T	Ethyl Acetate	50.000	47.492	5.0	101	0.00
38 T	Carbon Tetrachloride	50.000	47.650	4.7	102	0.00
39 T	Methylcyclohexane	50.000	48.174	3.7	102	0.00
40 TM	Benzene	50.000	47.569	4.9	102	0.00
41 T	Methacrylonitrile	50.000	50.810	-1.6	103	0.00
42 TM	1,2-Dichloroethane	50.000	47.020	6.0	102	0.00
43 T	Isopropyl Acetate	50.000	49.199	1.6	102	0.00
44 TM	Trichloroethene	50.000	47.486	5.0	102	0.00
45 C	1,2-Dichloropropane	50.000	47.169	5.7#	104	0.00
46 T	Dibromomethane	50.000	46.877	6.2	102	0.00
47 T	Bromodichloromethane	50.000	47.901	4.2	103	0.00
48 T	Methyl methacrylate	50.000	49.631	0.7	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061021\
 Data File : VU044142.D
 Acq On : 10 Jun 2021 09:46
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 01:17:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	973.402	2.7	98	0.00
50 S	Toluene-d8	50.000	46.677	6.6	98	0.00
51 T	4-Methyl-2-Pentanone	250.000	249.531	0.2	101	0.00
52 CM	Toluene	50.000	49.010	2.0#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	49.733	0.5	103	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.777	0.4	103	0.00
55 T	1,1,2-Trichloroethane	50.000	47.432	5.1	101	0.00
56 T	Ethyl methacrylate	50.000	52.023	-4.0	102	0.00
57 T	1,3-Dichloropropane	50.000	47.682	4.6	102	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	220.967	11.6	100	0.00
59 T	2-Hexanone	250.000	247.701	0.9	101	0.00
60 T	Dibromochloromethane	50.000	48.922	2.2	102	0.00
61 T	1,2-Dibromoethane	50.000	47.232	5.5	100	0.00
62 S	4-Bromofluorobenzene	50.000	47.395	5.2	101	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	47.597	4.8	102	0.00
65 PM	Chlorobenzene	50.000	47.350	5.3	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.820	4.4	101	0.00
67 C	Ethyl Benzene	50.000	49.769	0.5#	101	0.00
68 T	m/p-Xylenes	100.000	101.206	-1.2	101	0.00
69 T	o-Xylene	50.000	50.843	-1.7	102	0.00
70 T	Styrene	50.000	51.142	-2.3	101	0.00
71 P	Bromoform	50.000	48.454	3.1	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	52.145	-4.3	101	0.00
74 T	N-amyl acetate	50.000	50.356	-0.7	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.185	5.6	100	0.00
76 T	1,2,3-Trichloropropane	50.000	47.967	4.1	98	0.00
77 T	Bromobenzene	50.000	48.295	3.4	102	0.00
78 T	n-propylbenzene	50.000	50.632	-1.3	101	0.00
79 T	2-Chlorotoluene	50.000	50.075	-0.2	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.391	-4.8	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.038	-0.1	101	0.00
82 T	4-Chlorotoluene	50.000	50.960	-1.9	102	0.00
83 T	tert-Butylbenzene	50.000	51.068	-2.1	100	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.600	-5.2	99	0.00
85 T	sec-Butylbenzene	50.000	51.822	-3.6	102	0.00
86 T	p-Isopropyltoluene	50.000	53.300	-6.6	101	0.00
87 T	1,3-Dichlorobenzene	50.000	48.359	3.3	101	0.00
88 T	1,4-Dichlorobenzene	50.000	45.639	8.7	99	0.00
89 T	n-Butylbenzene	50.000	50.508	-1.0	101	0.00
90 T	Hexachloroethane	50.000	49.990	0.0	103	0.00
91 T	1,2-Dichlorobenzene	50.000	47.040	5.9	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.279	5.4	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.341	5.3	99	0.00
94 T	Hexachlorobutadiene	50.000	47.794	4.4	99	0.00
95 T	Naphthalene	50.000	44.132	11.7	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061021\
Data File : VU044142.D
Acq On : 10 Jun 2021 09:46
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Jun 11 01:17:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	46.943	6.1	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6