

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\U061021\
 Data File : U044170.D
 Acq On : 10 Jun 2021 20:52
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 01:22:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	50.000	46.688	6.6	96	0.00
3 P	Chloromethane	50.000	48.030	3.9	93	0.00
4 C	Vinyl Chloride	50.000	45.325	9.3#	96	0.00
5 T	Bromomethane	50.000	63.601	-27.2#	115	0.00
6 T	Chloroethane	50.000	50.284	-0.6	98	0.00
7 T	Trichlorofluoromethane	50.000	47.221	5.6	99	0.00
8 T	Diethyl Ether	50.000	46.871	6.3	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.307	7.4	98	0.00
10 T	Methyl Iodide	50.000	54.557	-9.1	103	0.00
11 T	Tert butyl alcohol	250.000	203.147	18.7	86	-0.01
12 CM	1,1-Dichloroethene	50.000	47.373	5.3#	100	0.00
13 T	Acrolein	250.000	177.255	29.1#	70	0.00
14 T	Allyl chloride	50.000	43.961	12.1	91	0.00
15 T	Acrylonitrile	250.000	229.966	8.0	92	0.00
16 T	Acetone	250.000	205.733	17.7	83	0.00
17 T	Carbon Disulfide	50.000	44.979	10.0	97	0.00
18 T	Methyl Acetate	50.000	46.450	7.1	94	0.00
19 T	Methyl tert-butyl Ether	50.000	49.991	0.0	99	0.00
20 T	Methylene Chloride	50.000	51.010	-2.0	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.358	3.3	100	0.00
22 T	Diisopropyl ether	50.000	48.024	4.0	95	0.00
23 T	Vinyl Acetate	250.000	240.086	4.0	94	0.00
24 P	1,1-Dichloroethane	50.000	46.542	6.9	98	0.00
25 T	2-Butanone	250.000	213.832	14.5	87	0.00
26 T	2,2-Dichloropropane	50.000	40.364	19.3	85	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.309	3.4	101	0.00
28 T	Bromochloromethane	50.000	47.676	4.6	99	0.00
29 T	Tetrahydrofuran	250.000	230.083	8.0	90	0.00
30 C	Chloroform	50.000	47.727	4.5#	100	0.00
31 T	Cyclohexane	50.000	44.745	10.5	95	0.00
32 T	1,1,1-Trichloroethane	50.000	48.073	3.9	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.750	6.5	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	47.186	5.6	97	0.00
36 T	1,1-Dichloropropene	50.000	45.818	8.4	97	0.00
37 T	Ethyl Acetate	50.000	44.103	11.8	92	0.00
38 T	Carbon Tetrachloride	50.000	47.625	4.8	100	0.00
39 T	Methylcyclohexane	50.000	45.247	9.5	94	0.00
40 TM	Benzene	50.000	47.229	5.5	99	0.00
41 T	Methacrylonitrile	50.000	46.745	6.5	92	0.00
42 TM	1,2-Dichloroethane	50.000	46.049	7.9	98	0.00
43 T	Isopropyl Acetate	50.000	45.975	8.0	94	0.00
44 TM	Trichloroethene	50.000	47.907	4.2	101	0.00
45 C	1,2-Dichloropropane	50.000	45.408	9.2#	98	0.00
46 T	Dibromomethane	50.000	46.884	6.2	100	0.00
47 T	Bromodichloromethane	50.000	47.275	5.5	99	0.00
48 T	Methyl methacrylate	50.000	45.821	8.4	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061021\
 Data File : VU044170.D
 Acq On : 10 Jun 2021 20:52
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 01:22:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	871.793	12.8	86	0.00
50 S	Toluene-d8	50.000	46.808	6.4	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	226.143	9.5	90	0.00
52 CM	Toluene	50.000	49.567	0.9#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	46.628	6.7	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.119	5.8	96	0.00
55 T	1,1,2-Trichloroethane	50.000	47.347	5.3	99	0.00
56 T	Ethyl methacrylate	50.000	49.294	1.4	95	0.00
57 T	1,3-Dichloropropane	50.000	47.599	4.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	237.736	4.9	106	0.00
59 T	2-Hexanone	250.000	220.470	11.8	88	0.00
60 T	Dibromochloromethane	50.000	48.593	2.8	100	0.00
61 T	1,2-Dibromoethane	50.000	47.978	4.0	100	0.00
62 S	4-Bromofluorobenzene	50.000	46.084	7.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64 T	Tetrachloroethene	50.000	47.844	4.3	99	0.00
65 PM	Chlorobenzene	50.000	47.437	5.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.662	2.7	100	0.00
67 C	Ethyl Benzene	50.000	49.687	0.6#	99	0.00
68 T	m/p-Xylenes	100.000	102.052	-2.1	99	0.00
69 T	o-Xylene	50.000	51.086	-2.2	100	0.00
70 T	Styrene	50.000	50.846	-1.7	98	0.00
71 P	Bromoform	50.000	48.376	3.2	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	51.364	-2.7	98	0.00
74 T	N-amyl acetate	50.000	47.828	4.3	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.528	8.9	95	0.00
76 T	1,2,3-Trichloropropane	50.000	45.310	9.4	91	0.00
77 T	Bromobenzene	50.000	48.398	3.2	100	0.00
78 T	n-propylbenzene	50.000	48.918	2.2	96	0.00
79 T	2-Chlorotoluene	50.000	48.296	3.4	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.683	-3.4	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.407	11.2	88	0.00
82 T	4-Chlorotoluene	50.000	48.674	2.7	96	0.00
83 T	tert-Butylbenzene	50.000	50.755	-1.5	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.956	-3.9	97	0.00
85 T	sec-Butylbenzene	50.000	50.366	-0.7	97	0.00
86 T	p-Isopropyltoluene	50.000	51.266	-2.5	96	0.00
87 T	1,3-Dichlorobenzene	50.000	48.224	3.6	99	0.00
88 T	1,4-Dichlorobenzene	50.000	45.669	8.7	97	0.00
89 T	n-Butylbenzene	50.000	47.548	4.9	94	0.00
90 T	Hexachloroethane	50.000	48.347	3.3	98	0.00
91 T	1,2-Dichlorobenzene	50.000	47.178	5.6	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	44.732	10.5	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.452	3.1	99	0.00
94 T	Hexachlorobutadiene	50.000	47.478	5.0	97	0.00
95 T	Naphthalene	50.000	46.577	6.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061021\
Data File : VU044170.D
Acq On : 10 Jun 2021 20:52
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Jun 11 01:22:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.957	0.1	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6