

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061022\
 Data File : VU049000.D
 Acq On : 10 Jun 2022 16:39
 Operator : SY/MD
 Sample : N3254-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0DF0

Quant Time: Jun 11 02:43:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 10 22:44:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	248771	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	239828	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	101299	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	77998	50.894	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.780%
7) Chloroethane-d5	1.893	69	57565	51.157	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.320%
11) 1,1-Dichloroethene-d2	2.559	63	105797	36.585	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.160%
21) 2-Butanone-d5	4.620	46	179734	106.967	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.970%
24) Chloroform-d	5.060	84	170097	52.028	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.060%
26) 1,2-Dichloroethane-d4	5.700	65	114730	54.167	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.340%
32) Benzene-d6	5.726	84	296159	49.800	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	6.690	67	106057	53.049	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.100%
41) Toluene-d8	7.896	98	259354	49.451	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.900%
43) trans-1,3-Dichloroprop...	8.179	79	40617	43.185	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.380%
47) 2-Hexanone-d5	8.632	63	120804	115.455	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	115.450%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	183644	54.998	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	102961	58.416	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	116.840%
Target Compounds						
13) Acetone	2.626	43	4196	3.174	ug/L	Qvalue 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061022\
Data File : VU049000.D
Acq On : 10 Jun 2022 16:39
Operator : SY/MD
Sample : N3254-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0DF0

Quant Time: Jun 11 02:43:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 10 22:44:31 2022
Response via : Initial Calibration





#13

Acetone

Concen: 3.174 ug/L

RT: 2.626 min Scan# 40

Delta R.T. 0.003 min

Lab File: VU049000.D


Acq: 10 Jun 2022 16:39

Instrument :

MSVOA_U

ClientSampleId :

C0DF0



Tgt Ion: 43 Resp: 4196

Ion Ratio Lower Upper

43 100

58 28.6 0.0 66.4

