

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061025\
 Data File : VU063374.D
 Acq On : 10 Jun 2025 10:36
 Operator : MD/SY
 Sample : VU0610WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK165

Quant Time: Jun 10 10:59:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060625WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 09 11:34:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	91046	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	96246	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	40145	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	45486	5.092	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.800%	
7) Chloroethane-d5	1.904	69	37941	5.427	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	108.600%	
11) 1,1-Dichloroethene-d2	2.557	65	16004	5.143	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	102.800%	
20) 2-Butanone-d5	4.615	46	105940	55.694	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.380%	
24) Chloroform-d	5.049	84	78594	5.617	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	112.400%	
26) 1,2-Dichloroethane-d4	5.692	65	41647	5.855	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	117.000%	
32) Benzene-d6	5.714	84	148709	5.127	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.600%	
36) 1,2-Dichloropropane-d6	6.679	67	52685	5.456	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.200%	
41) Toluene-d8	7.891	98	116340	4.721	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.400%	
43) trans-1,3-Dichloroprop...	8.174	79	16375	4.867	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	97.400%	
46) 2-Hexanone-d5	8.624	63	67271	45.519	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.040%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	43812	5.227	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	39352	5.854	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	117.000%	
Target Compounds						
					Qvalue	
35) Methylcyclohexane	6.682	83	11645	0.861	ug/L #	18
37) 1,2-Dichloropropane	6.679	63	5519	0.544	ug/L #	92
61) 1,2,3-Trichloropropane	11.807	75	5372	1.103	ug/L #	68
71) Naphthalene	14.122	128	1841	0.174	ug/L #	79
72) 1,2,3-Trichlorobenzene	14.338	180	1033	0.153	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061025\
Data File : VU063374.D
Acq On : 10 Jun 2025 10:36
Operator : MD/SY
Sample : VU0610WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK165

Quant Time: Jun 10 10:59:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060625WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 09 11:34:08 2025
Response via : Initial Calibration

