

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061119\
 Data File : VU032606.D
 Acq On : 11 Jun 2019 13:45
 Operator : JC/SP
 Sample : K3277-24
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 12 07:17:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 12 05:31:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	175852	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	167095	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	85047	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	57714	50.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.18%
7) Chloroethane-d5	1.68	69	46907	51.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.32%
11) 1,1-Dichloroethene-d2	2.27	63	86342	38.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.30%
21) 2-Butanone-d5	4.17	46	98612	103.28	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.28%
24) Chloroform-d	4.64	84	108424	50.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.90%
26) 1,2-Dichloroethane-d4	5.30	65	72428	54.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.06%
32) Benzene-d6	5.33	84	221406	51.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.08%
36) 1,2-Dichloropropane-d6	6.32	67	70185	51.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.20%
41) Toluene-d8	7.56	98	203956	49.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.18%
43) trans-1,3-Dichloropropene-	7.85	79	31804	48.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.72%
47) 2-Hexanone-d5	8.31	63	59425	90.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.07%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	100297	50.89	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.78%
64) 1,2-Dichlorobenzene-d4	11.85	152	85871	52.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061119\
Data File : VU032606.D
Acq On : 11 Jun 2019 13:45
Operator : JC/SP
Sample : K3277-24
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 12 07:17:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 12 05:31:42 2019
Response via : Initial Calibration

