

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061119\
 Data File : VU032621.D
 Acq On : 11 Jun 2019 19:38
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: Jun 12 08:05:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 12 07:28:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	169141	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	162272	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	80772	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	53098	48.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.78%
7) Chloroethane-d5	1.68	69	43795	49.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.32%
11) 1,1-Dichloroethene-d2	2.27	63	81302	37.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.68%
21) 2-Butanone-d5	4.17	46	94703	103.12	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.12%
24) Chloroform-d	4.64	84	101695	49.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.38%
26) 1,2-Dichloroethane-d4	5.30	65	69915	54.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.46%
32) Benzene-d6	5.33	84	211578	50.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.44%
36) 1,2-Dichloropropane-d6	6.32	67	66468	50.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.64%
41) Toluene-d8	7.56	98	193053	48.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.66%
43) trans-1,3-Dichloropropene-	7.85	79	29524	46.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.40%
47) 2-Hexanone-d5	8.31	63	56116	87.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.58%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	95609	49.95	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.90%
64) 1,2-Dichlorobenzene-d4	11.86	152	80144	51.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061119\
Data File : VU032621.D
Acq On : 11 Jun 2019 19:38
Operator : JC/SP
Sample : VIBLK40
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jun 12 08:05:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 12 07:28:55 2019
Response via : Initial Calibration

