

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061120\
 Data File : VU038779.D
 Acq On : 10 Jun 2020 16:30
 Operator : SY/MD
 Sample : VU0611WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK43

Quant Time: Jun 11 03:51:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 03:42:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	493960	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	480511	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	229787	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	123843	44.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.74%
7) Chloroethane-d5	1.93	69	109321	51.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.12%
11) 1,1-Dichloroethene-d2	2.59	63	210040	34.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.84%
21) 2-Butanone-d5	4.67	46	295753	110.15	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.15%
24) Chloroform-d	5.10	84	300809	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.10%
26) 1,2-Dichloroethane-d4	5.74	65	202567	51.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.72%
32) Benzene-d6	5.76	84	588070	50.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.62%
36) 1,2-Dichloropropane-d6	6.72	67	182926	51.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.56%
41) Toluene-d8	7.92	98	506599	48.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.86%
43) trans-1,3-Dichloropropene-	8.20	79	87183	48.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.08%
47) 2-Hexanone-d5	8.66	63	211290	111.28	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.28%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	275604	51.79	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.58%
64) 1,2-Dichlorobenzene-d4	12.21	152	213743	53.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061120\
Data File : VU038779.D
Acq On : 10 Jun 2020 16:30
Operator : SY/MD
Sample : VU0611WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK43

Quant Time: Jun 11 03:51:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 03:42:48 2020
Response via : Initial Calibration

