

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061120\
 Data File : VU038792.D
 Acq On : 10 Jun 2020 21:52
 Operator : SY/MD
 Sample : L2915-01 100X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9E04

Quant Time: Jun 11 05:05:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 03:42:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	549698	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	523594	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	248395	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	183165	59.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	119.28%
7) Chloroethane-d5	1.93	69	147369	61.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	123.70%
11) 1,1-Dichloroethene-d2	2.59	63	276481	41.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.62%
21) 2-Butanone-d5	4.66	46	360390	120.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.62%
24) Chloroform-d	5.10	84	365817	55.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.50%
26) 1,2-Dichloroethane-d4	5.74	65	252309	57.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.98%
32) Benzene-d6	5.76	84	737308	58.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.92%
36) 1,2-Dichloropropane-d6	6.72	67	225589	58.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	116.08%
41) Toluene-d8	7.92	98	627268	55.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.08%
43) trans-1,3-Dichloropropene-	8.20	79	109262	55.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	111.66%
47) 2-Hexanone-d5	8.65	63	255482	123.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	123.48%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	326452	56.30	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	112.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	255919	59.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.68%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
29) Cyclohexane	5.42	56	16047	2.679	ug/L	97
33) Benzene	5.81	78	6274334	403.535	ug/L	100
42) Toluene	7.99	91	447977	27.555	ug/L	99
52) Ethylbenzene	9.59	91	535814	29.863	ug/L	99
53) m,p-Xylene	9.71	106	30488	4.427	ug/L	100
54) o-xylene	10.11	106	44587	6.668	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061120\
Data File : VU038792.D
Acq On : 10 Jun 2020 21:52
Operator : SY/MD
Sample : L2915-01 100X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9E04

Quant Time: Jun 11 05:05:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 03:42:48 2020
Response via : Initial Calibration

