

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061120\
 Data File : VU038810.D
 Acq On : 11 Jun 2020 05:13
 Operator : SY/MD
 Sample : L2915-17 100X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9E50

Quant Time: Jun 11 08:03:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 07:12:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	575467	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	555764	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	268515	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	177871	55.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.64%
7) Chloroethane-d5	1.93	69	148739	59.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	119.26%
11) 1,1-Dichloroethene-d2	2.59	63	280366	40.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.02%
21) 2-Butanone-d5	4.66	46	370804	118.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.55%
24) Chloroform-d	5.10	84	366194	52.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.66%
26) 1,2-Dichloroethane-d4	5.74	65	250762	54.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.16%
32) Benzene-d6	5.76	84	743378	55.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.06%
36) 1,2-Dichloropropane-d6	6.72	67	228277	55.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.66%
41) Toluene-d8	7.92	98	639829	52.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.78%
43) trans-1,3-Dichloropropene-	8.20	79	106845	51.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.88%
47) 2-Hexanone-d5	8.65	63	267380	121.75	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	121.75%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	334209	54.30	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	264766	56.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.58%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
29) Cyclohexane	5.42	56	15738	2.475	ug/L	93
33) Benzene	5.81	78	6085397	368.729	ug/L	100
42) Toluene	7.99	91	347051	20.111	ug/L	100
52) Ethylbenzene	9.59	91	448509	23.550	ug/L	99
53) m,p-Xylene	9.71	106	12785	1.749	ug/L	92
54) o-xylene	10.11	106	28991	4.085	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061120\
Data File : VU038810.D
Acq On : 11 Jun 2020 05:13
Operator : SY/MD
Sample : L2915-17 100X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9E50

Quant Time: Jun 11 08:03:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 07:12:16 2020
Response via : Initial Calibration

