

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061120\
 Data File : VU038801.D
 Acq On : 11 Jun 2020 01:55
 Operator : SY/MD
 Sample : VU0611WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK44

Quant Time: Jun 11 07:30:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 07:12:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	555028	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	546476	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	257279	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	119106	38.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.82%
7) Chloroethane-d5	1.93	69	103108	42.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.72%
11) 1,1-Dichloroethene-d2	2.59	63	207833	30.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.50%
21) 2-Butanone-d5	4.66	46	317005	105.08	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.08%
24) Chloroform-d	5.10	84	297583	44.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.02%
26) 1,2-Dichloroethane-d4	5.74	65	203799	45.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.98%
32) Benzene-d6	5.76	84	590512	44.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.72%
36) 1,2-Dichloropropane-d6	6.72	67	183750	45.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.58%
41) Toluene-d8	7.92	98	509535	42.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.66%
43) trans-1,3-Dichloropropene-	8.20	79	87296	42.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.48%
47) 2-Hexanone-d5	8.65	63	222576	103.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.07%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	277659	45.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.76%
64) 1,2-Dichlorobenzene-d4	12.20	152	215757	48.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061120\
Data File : VU038801.D
Acq On : 11 Jun 2020 01:55
Operator : SY/MD
Sample : VU0611WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK44

Quant Time: Jun 11 07:30:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 07:12:16 2020
Response via : Initial Calibration

