

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061122\
 Data File : VU049058.D
 Acq On : 11 Jun 2022 17:33
 Operator : SY/MD
 Sample : N3277-04 20X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B87

Quant Time: Jun 13 01:48:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 01:35:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	150423	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	151325	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	66322	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	32416	4.872	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.400%
7) Chloroethane-d5	1.916	69	37495	5.168	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.400%
11) 1,1-Dichloroethene-d2	2.568	65	17419	5.015	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.400%
20) 2-Butanone-d5	4.636	46	178392	53.855	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.700%
24) Chloroform-d	5.067	84	108007	5.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	5.703	65	62160	5.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.729	84	186347	4.586	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
36) 1,2-Dichloropropane-d6	6.690	67	68026	4.913	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.899	98	149242	4.371	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	8.179	79	20838	4.479	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.636	63	119680	46.964	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.920%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	56461	4.636	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	61029	5.002	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds						
					Qvalue	
13) Acetone	2.662	43	9893	5.770	ug/L	95
34) Trichloroethene	6.546	95	11390	0.972	ug/L	92
47) Tetrachloroethene	8.552	164	4646	0.499	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061122\
Data File : VU049058.D
Acq On : 11 Jun 2022 17:33
Operator : SY/MD
Sample : N3277-04 20X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B87

Quant Time: Jun 13 01:48:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 01:35:15 2022
Response via : Initial Calibration

