

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061122\
 Data File : VU049061.D
 Acq On : 11 Jun 2022 18:44
 Operator : SY/MD
 Sample : N3277-08 200X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BA7

Quant Time: Jun 13 01:49:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 01:35:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	138200	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	138342	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	62819	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	31440	5.144	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.800%
7) Chloroethane-d5	1.913	69	35090	5.265	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.565	65	17013	5.331	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.600%
20) 2-Butanone-d5	4.636	46	182453	59.952	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.900%
24) Chloroform-d	5.064	84	102911	5.209	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
26) 1,2-Dichloroethane-d4	5.704	65	60817	5.562	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.200%
32) Benzene-d6	5.729	84	183798	4.948	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.694	67	64275	5.078	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.600%
41) Toluene-d8	7.900	98	143665	4.602	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%
43) trans-1,3-Dichloroprop...	8.182	79	19999	4.702	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.000%
46) 2-Hexanone-d5	8.636	63	127380	54.676	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.360%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	59420	5.337	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	63996	5.538	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.800%
Target Compounds						Qvalue
9) Trichlorofluoromethane	2.141	101	70976	4.277	ug/L	98
18) trans-1,2-Dichloroethene	3.360	96	1066	0.114	ug/L #	91
22) cis-1,2-Dichloroethene	4.668	96	2034	0.184	ug/L	94
34) Trichloroethene	6.543	95	3101007	289.582	ug/L	97
42) Toluene	7.973	91	12724	0.302	ug/L	95
47) Tetrachloroethene	8.555	164	55923	6.570	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061122\
Data File : VU049061.D
Acq On : 11 Jun 2022 18:44
Operator : SY/MD
Sample : N3277-08 200X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BA7

Quant Time: Jun 13 01:49:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 01:35:15 2022
Response via : Initial Calibration

