

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061122\
 Data File : VU049073.D
 Acq On : 11 Jun 2022 23:29
 Operator : SY/MD
 Sample : N3277-14 2000X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BB3

Quant Time: Jun 13 01:51:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 01:35:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	132558	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	135464	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	60812	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	29841	5.090	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.800%	
7) Chloroethane-d5	1.916	69	34531	5.401	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	108.000%	
11) 1,1-Dichloroethene-d2	2.568	65	14799	4.835	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	96.600%	
20) 2-Butanone-d5	4.639	46	161667	55.383	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	110.760%	
24) Chloroform-d	5.063	84	98734	5.211	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.200%	
26) 1,2-Dichloroethane-d4	5.703	65	59981	5.720	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	114.400%	
32) Benzene-d6	5.729	84	174074	4.786	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.690	67	61966	4.999	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.000%	
41) Toluene-d8	7.899	98	133318	4.362	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.200%	
43) trans-1,3-Dichloroprop...	8.179	79	20257	4.864	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	97.200%	
46) 2-Hexanone-d5	8.639	63	123243	54.024	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	108.040%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	57387	5.264	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.200%	
66) 1,2-Dichlorobenzene-d4	12.192	152	59446	5.314	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.200%	
Target Compounds						
						Qvalue
9) Trichlorofluoromethane	2.141	101	5035	0.316	ug/L	89
34) Trichloroethene	6.542	95	168481	16.068	ug/L	95
47) Tetrachloroethene	8.552	164	13794	1.655	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061122\
Data File : VU049073.D
Acq On : 11 Jun 2022 23:29
Operator : SY/MD
Sample : N3277-14 2000X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BB3

Quant Time: Jun 13 01:51:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 01:35:15 2022
Response via : Initial Calibration

