

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061219\
 Data File : VU032638.D
 Acq On : 12 Jun 2019 09:58
 Operator : JC/SP
 Sample : VU0612WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK58

Quant Time: Jun 13 05:11:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 13 05:01:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	169531	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	158832	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	79969	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	50517	45.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.86%
7) Chloroethane-d5	1.68	69	41820	47.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.62%
11) 1,1-Dichloroethene-d2	2.27	63	77212	35.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.70%
21) 2-Butanone-d5	4.17	46	86696	94.19	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.19%
24) Chloroform-d	4.64	84	100384	48.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.88%
26) 1,2-Dichloroethane-d4	5.30	65	67030	51.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.74%
32) Benzene-d6	5.33	84	204139	49.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.02%
36) 1,2-Dichloropropane-d6	6.32	67	66239	51.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.46%
41) Toluene-d8	7.56	98	187157	47.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.74%
43) trans-1,3-Dichloropropene-	7.85	79	28768	46.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.98%
47) 2-Hexanone-d5	8.31	63	42825	68.28	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	68.28%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	91466	48.82	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.64%
64) 1,2-Dichlorobenzene-d4	11.86	152	77189	50.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.78%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061219\
Data File : VU032638.D
Acq On : 12 Jun 2019 09:58
Operator : JC/SP
Sample : VU0612WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK58

Quant Time: Jun 13 05:11:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 13 05:01:06 2019
Response via : Initial Calibration

