

Data File : VU038816.D
Acq On : 11 Jun 2020 11:45
Operator : SY/MD
Sample : L2910-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 12 01:17:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 12 01:02:05 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	328802	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	331797	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	157927	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	92435	4.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.80%
7) Chloroethane-d5	1.93	69	78228	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.59	63	147868	3.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.60%
20) 2-Butanone-d5	4.67	46	253367	39.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.38%
24) Chloroform-d	5.10	84	187347	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.74	65	114489	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.76	84	382719	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	6.72	67	117571	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	7.92	98	322400	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
43) trans-1,3-Dichloropropene-	8.20	79	48414	4.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.40%
46) 2-Hexanone-d5	8.66	63	287576	53.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.32%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	103579	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	134783	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds					Qvalue
16) Methylene chloride	3.08	84	3167	0.138 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU038816.D
Acq On : 11 Jun 2020 11:45
Operator : SY/MD
Sample : L2910-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 12 01:17:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 12 01:02:05 2020
Response via : Initial Calibration

