

Data File : VU038822.D
Acq On : 11 Jun 2020 14:15
Operator : SY/MD
Sample : L2933-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH6

Quant Time: Jun 12 01:50:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 12 01:02:05 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	283166	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	282643	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	133314	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	67234	3.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.40%
7) Chloroethane-d5	1.93	69	66437	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.59	63	116272	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	4.67	46	315279	56.63	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.26%
24) Chloroform-d	5.10	84	179210	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
26) 1,2-Dichloroethane-d4	5.74	65	108102	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.60%
32) Benzene-d6	5.76	84	345285	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.72	67	107193	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.80%
41) Toluene-d8	7.92	98	284667	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	8.20	79	41648	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
46) 2-Hexanone-d5	8.65	63	256252	55.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.22%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	95380	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	121822	5.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.00%

Target Compounds					Qvalue
13) Acetone	2.67	43	49251	13.518 ug/L	100
14) Carbon disulfide	2.82	76	5367	0.091 ug/L #	86
25) Chloroform	5.12	83	51736	1.514 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU038822.D
Acq On : 11 Jun 2020 14:15
Operator : SY/MD
Sample : L2933-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH6

Quant Time: Jun 12 01:50:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 12 01:02:05 2020
Response via : Initial Calibration

