


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU061220\
 Data File : VU038813.D
 Acq On : 11 Jun 2020 08:45
 Operator : SY/MD
 Sample : BFB11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB11

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Jun 15 02:08:46 2020



AutoFind: Scans 2894, 2895, 2896; Background Corrected with Scan 2880

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	28712	PASS
75	95	30	80	53.0	80184	PASS
95	95	100	100	100.0	151424	PASS
96	95	5	9	6.8	10260	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	78.7	119133	PASS
175	174	5	9	7.4	8783	PASS
176	174	95	101	96.0	114394	PASS
177	176	5	9	6.5	7485	PASS

m/z	Abundance
39.90	215.0
43.90	552.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

m/z	Abundance
36.00	1312.0
37.00	7084.0
38.00	6485.0
39.00	2665.0
39.90	356.0
42.80	106.0
44.00	1446.0
45.00	1296.0
46.00	159.0
47.00	1920.0
47.90	885.0
49.00	6133.0
50.00	28616.0
51.00	8751.0
52.10	300.0
55.00	611.0
56.10	2002.0
57.00	3926.0
57.90	125.0
59.90	1229.0
61.10	7296.0
62.00	8233.0
63.10	6158.0
64.00	945.0
64.90	1099.0
66.00	122.0
67.10	445.0
68.00	15015.0
69.00	15651.0
70.00	1327.0
72.00	739.0
73.00	7512.0
74.10	27296.0
75.00	80200.0
76.00	6926.0
77.00	958.0
78.00	734.0
78.90	5081.0
79.90	1737.0
80.90	5535.0
81.90	1345.0
82.90	130.0
85.80	251.0
87.00	6814.0
88.00	5836.0
91.00	636.0
92.00	4575.0
93.00	6682.0
94.00	18760.0
95.10	149952.0
96.10	10467.0
97.00	391.0
103.90	846.0
105.00	286.0
105.90	718.0
107.00	133.0
109.70	103.0
110.80	160.0
112.90	236.0
114.80	172.0
115.90	637.0
117.00	1134.0
118.00	758.0
118.90	824.0
122.90	163.0
123.90	147.0
125.90	105.0
127.90	488.0
129.00	406.0
130.00	449.0
130.80	303.0
135.00	205.0
137.00	268.0
141.00	1454.0
141.90	216.0
143.00	1881.0
144.10	144.0
144.90	399.0
145.90	275.0
147.00	187.0
147.80	426.0
154.00	129.0
155.00	453.0
156.90	228.0
159.00	184.0
160.90	116.0
172.00	1599.0
174.00	116552.0
175.10	8846.0
176.00	111696.0
177.00	7418.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

177.90

274.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

m/z	Abundance
36.10	1361.0
37.00	7476.0
38.00	6896.0
39.00	2686.0
39.90	364.0
44.00	1385.0
44.90	1404.0
45.90	102.0
47.10	1918.0
47.90	863.0
49.00	6308.0
50.00	29600.0
51.00	9055.0
52.00	299.0
53.10	100.0
55.00	592.0
56.00	2127.0
57.00	4064.0
57.90	142.0
58.20	130.0
60.00	1380.0
61.10	7561.0
62.00	8438.0
63.10	6290.0
64.00	825.0
65.00	1004.0
66.00	128.0
68.00	15995.0
69.00	16042.0
70.00	1469.0
72.00	714.0
73.00	7716.0
74.10	27976.0
75.00	82624.0
76.00	6750.0
77.00	1071.0
78.10	854.0
78.90	5417.0
79.90	1628.0
80.90	5765.0
81.90	1287.0
82.90	151.0
85.80	320.0
87.00	6944.0
88.00	6050.0
91.00	679.0
92.00	4545.0
93.00	7129.0
94.10	20048.0
95.10	156672.0
96.10	10592.0
97.00	397.0
104.00	819.0
105.00	341.0
106.00	737.0
106.90	142.0
109.80	135.0
110.80	168.0
112.90	207.0
114.80	140.0
115.90	646.0
117.00	1246.0
118.00	654.0
119.00	807.0
122.90	119.0
123.90	126.0
125.90	125.0
128.00	568.0
129.00	365.0
130.00	442.0
130.80	299.0
135.00	242.0
137.00	228.0
141.00	1558.0
141.90	232.0
143.00	1893.0
144.00	143.0
144.90	401.0
145.90	241.0
147.10	144.0
147.90	394.0
153.90	111.0
155.00	386.0
156.00	158.0
157.00	228.0
158.90	182.0
160.90	137.0
172.00	1696.0
174.00	122856.0
175.10	8976.0
176.00	117816.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

177.007686.0

177.90309.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

m/z	Abundance
36.10	1194.0
37.00	7242.0
38.00	6591.0
39.00	2421.0
39.90	338.0
42.80	104.0
44.00	1317.0
44.90	1397.0
47.10	1569.0
47.90	864.0
49.00	5710.0
50.00	27920.0
51.00	8580.0
51.90	284.0
55.00	467.0
56.00	2157.0
57.10	3911.0
57.90	167.0
60.00	1602.0
61.00	6963.0
62.00	7865.0
63.00	5868.0
64.00	663.0
65.00	876.0
66.10	126.0
68.00	15261.0
69.00	15060.0
70.00	1333.0
72.00	728.0
73.00	7002.0
74.00	26120.0
75.00	77728.0
76.00	6451.0
77.00	996.0
78.90	5091.0
79.90	1409.0
80.90	5487.0
82.00	1296.0
82.90	160.0
85.80	275.0
87.00	6401.0
88.00	5617.0
88.80	159.0
91.00	622.0
92.00	4066.0
93.00	6943.0
94.10	19216.0
95.10	147648.0
96.10	9723.0
97.10	307.0
104.00	700.0
104.90	350.0
106.00	749.0
106.80	116.0
109.80	128.0
110.80	164.0
113.00	176.0
114.90	110.0
115.90	614.0
117.00	1199.0
118.00	509.0
119.00	757.0
126.00	110.0
128.00	557.0
129.10	303.0
129.90	432.0
130.80	264.0
134.90	273.0
137.00	189.0
141.00	1637.0
141.90	190.0
142.90	1894.0
144.00	109.0
144.90	368.0
145.90	199.0
147.90	316.0
149.80	126.0
155.00	336.0
156.00	180.0
157.00	223.0
158.90	150.0
161.00	160.0
172.00	1530.0
174.00	117992.0
175.10	8527.0
176.00	113672.0
177.00	7352.0
177.90	256.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11