

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061222\
 Data File : VU049090.D
 Acq On : 12 Jun 2022 14:49
 Operator : SY/MD
 Sample : N3255-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0DK4

Quant Time: Jun 13 05:04:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jun 13 05:01:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	241955	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	234420	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	98481	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	67955	45.590	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.180%
7) Chloroethane-d5	1.899	69	52803	48.247	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.500%
11) 1,1-Dichloroethene-d2	2.562	63	95899	34.096	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.200%
21) 2-Butanone-d5	4.616	46	154157	94.330	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.330%
24) Chloroform-d	5.063	84	157969	49.680	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.360%
26) 1,2-Dichloroethane-d4	5.700	65	109732	53.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.540%
32) Benzene-d6	5.726	84	293311	50.459	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.920%
36) 1,2-Dichloropropane-d6	6.690	67	97830	50.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.120%
41) Toluene-d8	7.896	98	237904	46.407	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.820%
43) trans-1,3-Dichloroprop...	8.179	79	38740	42.140	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.280%
47) 2-Hexanone-d5	8.632	63	85997	84.085	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.090%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	156460	47.938	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.880%
66) 1,2-Dichlorobenzene-d4	12.192	152	94497	55.148	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061222\
Data File : VU049090.D
Acq On : 12 Jun 2022 14:49
Operator : SY/MD
Sample : N3255-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0DK4

Quant Time: Jun 13 05:04:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jun 13 05:01:30 2022
Response via : Initial Calibration

