

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
 Data File : VU054491.D
 Acq On : 12 Jun 2023 17:11
 Operator : MD/SY
 Sample : 03104-19
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONR2

Quant Time: Jun 13 02:31:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 13 02:28:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	289377	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	274037	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	134399	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	51906	2.818	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	56.400%
7) Chloroethane-d5	1.914	69	62817	3.643	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.800%
11) 1,1-Dichloroethene-d2	2.563	65	26029	3.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.200%
20) 2-Butanone-d5	4.634	46	264784	51.138	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.280%
24) Chloroform-d	5.058	84	152268	3.956	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
26) 1,2-Dichloroethane-d4	5.698	65	79471	4.235	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.800%
32) Benzene-d6	5.724	84	286857	3.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.400%
36) 1,2-Dichloropropane-d6	6.685	67	104383	4.103	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.000%
41) Toluene-d8	7.894	98	238710	3.531	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.600%
43) trans-1,3-Dichloroprop...	8.177	79	33442	3.780	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.600%
46) 2-Hexanone-d5	8.631	63	166687	44.911	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.820%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	86698	4.570	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	99199	4.376	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.600%
Target Compounds						Qvalue
8) Chloroethane	1.933	64	6099	0.319	ug/L	97
9) Trichlorofluoromethane	2.139	101	9445	0.241	ug/L	100
12) 1,1-Dichloroethene	2.579	96	209413	9.977	ug/L	92
19) 1,1-Dichloroethane	3.866	63	2947056	65.714	ug/L	99
22) cis-1,2-Dichloroethene	4.663	96	21418	0.868	ug/L	96
25) Chloroform	5.084	83	20798	0.470	ug/L	96
27) 1,2-Dichloroethane	5.795	62	6085	0.235	ug/L	97
29) 1,1,1-Trichloroethane	5.312	97	7078688	190.494	ug/L	99
34) Trichloroethene	6.537	95	228443	8.772	ug/L	96
45) 1,1,2-Trichloroethane	8.396	97	3779	0.223	ug/L	91
47) Tetrachloroethene	8.550	164	4611	0.231	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
Data File : VU054491.D
Acq On : 12 Jun 2023 17:11
Operator : MD/SY
Sample : 03104-19
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONR2

Quant Time: Jun 13 02:31:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 13 02:28:29 2023
Response via : Initial Calibration

