

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
 Data File : VU054482.D
 Acq On : 12 Jun 2023 13:33
 Operator : MD/SY
 Sample : 03104-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONR8

Quant Time: Jun 13 02:29:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 13 02:28:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	299910	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	286196	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	140096	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	59187	3.101	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.000%
7) Chloroethane-d5	1.914	69	66579	3.726	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.600%
11) 1,1-Dichloroethene-d2	2.567	65	29474	3.458	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.200%
20) 2-Butanone-d5	4.637	46	273550	50.975	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.960%
24) Chloroform-d	5.058	84	165224	4.141	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.800%
26) 1,2-Dichloroethane-d4	5.698	65	86164	4.430	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
32) Benzene-d6	5.724	84	316141	3.974	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
36) 1,2-Dichloropropane-d6	6.689	67	113905	4.287	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.800%
41) Toluene-d8	7.898	98	264594	3.747	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.000%
43) trans-1,3-Dichloroprop...	8.180	79	36559	3.957	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.200%
46) 2-Hexanone-d5	8.631	63	168057	43.356	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.720%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	92539	4.671	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	104288	4.413	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
Data File : VU054482.D
Acq On : 12 Jun 2023 13:33
Operator : MD/SY
Sample : 03104-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONR8

Quant Time: Jun 13 02:29:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 13 02:28:29 2023
Response via : Initial Calibration

