

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
 Data File : VU054483.D
 Acq On : 12 Jun 2023 13:57
 Operator : MD/SY
 Sample : 03104-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONQ7

Quant Time: Jun 13 02:29:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 13 02:28:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	290859	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	282633	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	136622	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	57785	3.122	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.400%
7) Chloroethane-d5	1.911	69	64613	3.728	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.600%
11) 1,1-Dichloroethene-d2	2.563	65	27954	3.381	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	4.637	46	269141	51.714	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.420%
24) Chloroform-d	5.059	84	162800	4.208	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
26) 1,2-Dichloroethane-d4	5.698	65	82495	4.374	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
32) Benzene-d6	5.724	84	306440	3.900	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.000%
36) 1,2-Dichloropropane-d6	6.689	67	112051	4.271	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.894	98	259021	3.715	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.200%
43) trans-1,3-Dichloroprop...	8.177	79	36118	3.958	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.200%
46) 2-Hexanone-d5	8.631	63	166435	43.479	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.960%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	87952	4.495	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	104165	4.520	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
Data File : VU054483.D
Acq On : 12 Jun 2023 13:57
Operator : MD/SY
Sample : 03104-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONQ7

Quant Time: Jun 13 02:29:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 13 02:28:29 2023
Response via : Initial Calibration

