

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
 Data File : VU054486.D
 Acq On : 12 Jun 2023 15:10
 Operator : MD/SY
 Sample : 03104-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONR1

Quant Time: Jun 13 02:30:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 13 02:28:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	289973	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	278814	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	134972	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	58920	3.193	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.800%
7) Chloroethane-d5	1.911	69	67306	3.895	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.563	65	28421	3.448	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.000%
20) 2-Butanone-d5	4.634	46	281049	54.167	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.340%
24) Chloroform-d	5.058	84	172009	4.459	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	5.698	65	89903	4.781	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
32) Benzene-d6	5.724	84	322547	4.162	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	6.685	67	118114	4.563	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.200%
41) Toluene-d8	7.894	98	270801	3.937	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	8.177	79	38599	4.288	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.800%
46) 2-Hexanone-d5	8.631	63	184045	48.738	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.480%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	94470	4.894	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	111216	4.885	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%
Target Compounds						
					Qvalue	
29) 1,1,1-Trichloroethane	5.313	97	9215	0.244	ug/L	97
34) Trichloroethene	6.541	95	18939	0.715	ug/L	97
47) Tetrachloroethene	8.550	164	18408	0.908	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061223\
Data File : VU054486.D
Acq On : 12 Jun 2023 15:10
Operator : MD/SY
Sample : 03104-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONR1

Quant Time: Jun 13 02:30:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 13 02:28:29 2023
Response via : Initial Calibration





#29

1,1,1-Trichloroethane

Concen: 0.244 ug/L

RT: 5.313 min Scan# 11

Delta R.T. 0.000 min

Lab File: VU054486.D

Acq: 12 Jun 2023 15:10

Instrument :

MSVOA_U

ClientSampleId :

CONR1





Tgt Ion: 97 Resp: 9215

Ion Ratio Lower Upper

97 100

99 60.8 50.8 76.2

61 43.7 36.1 54.1



#34

Trichloroethene

Concen: 0.715 ug/L

RT: 6.541 min Scan# 1633

Delta R.T. 0.000 min

Lab File: VU054486.D

Acq: 12 Jun 2023 15:10

Tgt Ion: 95 Resp: 18939




Ion Ratio Lower Upper


95 100

97 65.5 46.8 87.0

132 94.6 68.4 127.0

130 101.0 68.7 127.5





#47
 Tetrachloroethene
 Concen: 0.908 ug/L
 RT: 8.550 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VU054486.D
 Acq: 12 Jun 2023 15:10

Instrument :
 MSVOA_U
 ClientSampleId :
 CONR1

Tgt Ion:164	Resp:	18408
Ion	Ratio	Lower Upper
164	100	
129	87.5	67.7 125.7
131	84.4	64.6 120.0
166	119.2	88.1 163.7

