

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061318\
 Data File : VU024523.D
 Acq On : 13 Jun 2018 21:14
 Operator : MD/SY
 Sample : J3461-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-48(0-5)

Quant Time: Jun 14 03:39:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	180919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	309533	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	299710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	175851	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	138987	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
35) Dibromofluoromethane	4.89	113	125883	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
50) Toluene-d8	7.57	98	455692	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	10.31	95	180144	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
Target Compounds						
16) Acetone	2.33	43	16740	10.549	ug/l	98
18) Methyl Acetate	2.62	43	16590	4.876	ug/l	97
20) Methylene Chloride	2.71	84	3391	1.362	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061318\
Data File : VU024523.D
Acq On : 13 Jun 2018 21:14
Operator : MD/SY
Sample : J3461-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-48(0-5)

Quant Time: Jun 14 03:39:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

