

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061318\
 Data File : VU024524.D
 Acq On : 13 Jun 2018 21:37
 Operator : MD/SY
 Sample : J3461-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-47(0-5)

Quant Time: Jun 14 03:40:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	189504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	326976	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	314817	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	185120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	146859	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
35) Dibromofluoromethane	4.89	113	129352	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
50) Toluene-d8	7.57	98	477086	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	10.31	95	190046	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
Target Compounds						
16) Acetone	2.33	43	17604	10.591	ug/l	93
18) Methyl Acetate	2.62	43	15431	4.330	ug/l	96
20) Methylene Chloride	2.71	84	3640	1.396	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061318\
Data File : VU024524.D
Acq On : 13 Jun 2018 21:37
Operator : MD/SY
Sample : J3461-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-47(0-5)

Quant Time: Jun 14 03:40:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

