

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061320\
 Data File : VU038845.D
 Acq On : 12 Jun 2020 12:41
 Operator : SY/MD
 Sample : VU0613WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK09

Quant Time: Jun 13 01:05:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 13 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	317182	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	308054	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	145501	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	72846	4.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.20%
7) Chloroethane-d5	1.93	69	58735	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.60%
11) 1,1-Dichloroethene-d2	2.59	63	122964	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	4.69	46	240270	44.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.56%
24) Chloroform-d	5.10	84	157526	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
26) 1,2-Dichloroethane-d4	5.74	65	92584	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.60%
32) Benzene-d6	5.76	84	320361	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
36) 1,2-Dichloropropane-d6	6.72	67	96898	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.92	98	278668	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	8.20	79	39933	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	8.66	63	203125	47.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.60%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	80591	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	106998	4.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.40%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	2796	0.109 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061320\
Data File : VU038845.D
Acq On : 12 Jun 2020 12:41
Operator : SY/MD
Sample : VU0613WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK09

Quant Time: Jun 13 01:05:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 13 00:56:28 2020
Response via : Initial Calibration

