

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061320\
 Data File : VU038851.D
 Acq On : 12 Jun 2020 15:09
 Operator : SY/MD
 Sample : L2957-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 13 01:44:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 13 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	275341	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	274457	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	125180	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	75089	5.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.00%
7) Chloroethane-d5	1.93	69	60465	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.59	63	120737	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	4.68	46	249884	53.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.10%
24) Chloroform-d	5.10	84	163210	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.74	65	97267	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.60%
32) Benzene-d6	5.76	84	328551	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.72	67	99918	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.92	98	277241	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
43) trans-1,3-Dichloropropene-	8.20	79	40797	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.65	63	208040	54.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.90%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	86473	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.00%
64) 1,2-Dichlorobenzene-d4	12.20	152	108872	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%


Target Compounds					Ovalue
16) Methylene chloride	3.08	84	5475	0.247 ug/L	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061320\
Data File : VU038851.D
Acq On : 12 Jun 2020 15:09
Operator : SY/MD
Sample : L2957-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 13 01:44:40 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 13 00:56:28 2020
Response via : Initial Calibration





#16

Methylene chloride

Concen: 0.247 ug/L

RT: 3.08 min Scan# 540

Delta R.T. 0.00 min

Lab File: VU038851.D

Acq: 12 Jun 2020 15:09

Instrument :

MSVOA_U

ClientSampleId :

VHBLK01

Tot Ion: 84 Resp: 5475

Ion Ratio Lower Upper

84 100

86 76.6 43.0 80.0

49 114.3 85.0 157.8

