

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061320\
 Data File : VU038853.D
 Acq On : 12 Jun 2020 15:55
 Operator : SY/MD
 Sample : L2985-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT77

Quant Time: Jun 13 01:52:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 13 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	268738	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	268223	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	125515	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	67554	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.93	69	64632	5.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.80%
11) 1,1-Dichloroethene-d2	2.59	63	113292	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	4.68	46	312776	68.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	136.08%#
24) Chloroform-d	5.10	84	171171	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
26) 1,2-Dichloroethane-d4	5.74	65	102750	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
32) Benzene-d6	5.76	84	332218	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
36) 1,2-Dichloropropane-d6	6.72	67	106284	5.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.40%
41) Toluene-d8	7.92	98	260822	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	8.20	79	38363	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	8.65	63	262803	71.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	142.06%#
57) 1,1,2,2-Tetrachloroethane-	10.77	84	98572	6.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.20%#
64) 1,2-Dichlorobenzene-d4	12.20	152	111312	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061320\
Data File : VU038853.D
Acq On : 12 Jun 2020 15:55
Operator : SY/MD
Sample : L2985-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFT77

Quant Time: Jun 13 01:52:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 13 00:56:28 2020
Response via : Initial Calibration

