

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061320\
 Data File : VU038879.D
 Acq On : 13 Jun 2020 03:36
 Operator : SY/MD
 Sample : L2985-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT74

Quant Time: Jun 13 04:24:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 13 02:52:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	254390	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	256737	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	117657	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	60248	4.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.00%
7) Chloroethane-d5	1.93	69	62963	5.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.00%
11) 1,1-Dichloroethene-d2	2.59	63	117993	3.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.80%
20) 2-Butanone-d5	4.68	46	308591	70.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	141.84%#
24) Chloroform-d	5.10	84	172375	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
26) 1,2-Dichloroethane-d4	5.74	65	104179	6.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	120.00%
32) Benzene-d6	5.76	84	320768	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.00%
36) 1,2-Dichloropropane-d6	6.72	67	104216	5.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	116.20%
41) Toluene-d8	7.92	98	252693	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloropropene-	8.20	79	35399	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.66	63	256440	72.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	144.82%#
57) 1,1,2,2-Tetrachloroethane-	10.77	84	94756	6.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.80%#
64) 1,2-Dichlorobenzene-d4	12.20	152	109811	5.74	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.80%

Target Compounds


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	1.39	85	19813	0.799	ug/L	99
9) Trichlorofluoromethane	2.16	101	116035	3.600	ug/L	100
10) 1,1,2-Trichloro-1,2,2-trif	2.60	101	5832	0.316	ug/L	93
12) 1,1-Dichloroethene	2.60	96	16386	0.920	ug/L #	1
19) 1,1-Dichloroethane	3.91	63	6559	0.191	ug/L #	94
22) cis-1,2-Dichloroethene	4.71	96	91153	4.575	ug/L	96
29) 1,1,1-Trichloroethane	5.35	97	4664	0.147	ug/L	93
31) Carbon tetrachloride	5.56	117	3707	0.139	ug/L	95
34) Trichloroethene	6.57	95	679378	33.535	ug/L	97
47) Tetrachloroethene	8.57	164	459656	29.473	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061320\
Data File : VU038879.D
Acq On : 13 Jun 2020 03:36
Operator : SY/MD
Sample : L2985-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFT74

Quant Time: Jun 13 04:24:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 13 02:52:53 2020
Response via : Initial Calibration





#2

Dichlorodifluoromethane

Concen: 0.799 ug/L

RT: 1.39 min Scan# 17

Delta R.T. -0.00 min

Lab File: VU038879.D

Acq: 13 Jun 2020 03:36

Instrument :

MSVOA_U

ClientSampleId :


BFT74

Tot Ion: 85 Resp: 19813

Ion Ratio Lower Upper




85 100

87 31.6 25.9 38.9



Abundance Ion 85.00 (84.70 to 85.70): VU038863.D (-9) (-)

Ion 87.00 (86.70 to 87.70): VU038879.D (-1) (-)



#9

Trichlorofluoromethane

Concen: 3.600 ug/L

RT: 2.16 min Scan# 254

Delta R.T. -0.00 min

Lab File: VU038879.D


Acq: 13 Jun 2020 03:36

Tgt Ion: 101 Resp: 116035

Ion Ratio Lower Upper



101 100


103 64.4 51.5 77.3



Abundance Ion 101.00 (100.70 to 101.70): VU038863.D (-243) (-)

Ion 103.00 (102.70 to 103.70): VU038879.D (-161) (-)





#10

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 0.316 ug/L

RT: 2.60 min Scan# 392

Delta R.T. -0.01 min

Lab File: VU038879.D

Acq: 13 Jun 2020 03:36

Instrument :

MSVOA_U

ClientSampled :

BFT74


Tot Ion:101 Resp: 5832

Ion Ratio Lower Upper

101 100

85 37.9 34.5 51.7




151 65.0 56.5 84.7



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VU0

Ion 151.00 (150.70 to 151.70): V



#12

1,1-Dichloroethene

Concen: 0.920 ug/L

RT: 2.60 min Scan# 393

Delta R.T. -0.00 min

Lab File: VU038879.D

Acq: 13 Jun 2020 03:36


Tgt Ion: 96 Resp: 16386

Ion Ratio Lower Upper

96 100

61 213.8 120.2 223.2



63 400.1 90.1 167.3#




Abundance Ion 96.00 (95.70 to 96.70): VU0

Ion 61.05 (60.75 to 61.75): VU0

Ion 63.00 (62.70 to 63.70): VU0









#19
 1,1-Dichloroethane
 Concen: 0.191 ug/L
 RT: 3.91 min Scan# 799
 Delta R.T. -0.00 min
 Lab File: VU038879.D
 Acq: 13 Jun 2020 03:36




Instrument :
 MSVOA_U
 ClientSampleId :
 BFT74


Tot Ion: 63 Resp: 6559
 Ion Ratio Lower Upper
 63 100
 65 33.8 22.3 41.5
 83 18.9 10.1 18.9#



#22
 cis-1,2-Dichloroethene
 Concen: 4.575 ug/L
 RT: 4.71 min Scan# 1047
 Delta R.T. -0.00 min
 Lab File: VU038879.D
 Acq: 13 Jun 2020 03:36

Tgt Ion: 96 Resp: 91153
 Ion Ratio Lower Upper
 96 100
 61 133.2 89.7 166.7
 68 0.0 0.0 0.0









#29
 1,1,1-Trichloroethane
 Concen: 0.147 ug/L
 RT: 5.35 min Scan# 1248
 Delta R.T. -0.00 min
 Lab File: VU038879.D
 Acq: 13 Jun 2020 03:36

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT74

Tot Ion: 97 Resp: 4664
 Ion Ratio Lower Upper
 97 100
 99 59.0 52.1 78.1
 61 48.2 35.9 53.9




Abundance Ion 96.95 (96.65 to 97.65): VU0
 Ion 99.05 (98.75 to 99.75): VU0
 Ion 61.05 (60.75 to 61.75): VU0






#31
 Carbon tetrachloride
 Concen: 0.139 ug/L
 RT: 5.56 min Scan# 1312
 Delta R.T. -0.00 min
 Lab File: VU038879.D
 Acq: 13 Jun 2020 03:36

Tgt Ion: 117 Resp: 3707
 Ion Ratio Lower Upper
 117 100
 119 89.9 76.2 114.2



Abundance Ion 116.90 (116.60 to 117.60): V
 Ion 118.90 (118.60 to 119.60): V









#34
 Trichloroethene
 Concen: 33.535 ug/L
 RT: 6.57 min Scan# 1627
 Delta R.T. -0.00 min
 Lab File: VU038879.D
 Acq: 13 Jun 2020 03:36

Instrument :
 MSVOA_U
 ClientSampled :
 BFT74

Tot Ion: 95 Resp: 679378
 Ion Ratio Lower Upper
 95 100
 97 65.3 46.1 85.5
 132 95.3 67.5 125.3
 130 100.4 66.7 123.9



#47
 Tetrachloroethene
 Concen: 29.473 ug/L
 RT: 8.57 min Scan# 2250
 Delta R.T. 0.00 min
 Lab File: VU038879.D
 Acq: 13 Jun 2020 03:36

Tgt Ion: 164 Resp: 459656
 Ion Ratio Lower Upper
 164 100
 129 97.2 67.3 124.9
 131 92.7 65.0 120.6
 166 129.0 91.0 169.0

