

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061419\
 Data File : VU032740.D
 Acq On : 14 Jun 2019 22:35
 Operator : JC/SP
 Sample : K3383-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0XK0

Quant Time: Jun 17 09:00:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 23:49:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	99860	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	95365	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	47331	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	27817	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	1.67	69	24888	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	2.27	63	45129	3.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.60%
20) 2-Butanone-d5	4.18	46	102412	56.52	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.04%
24) Chloroform-d	4.64	84	61173	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
26) 1,2-Dichloroethane-d4	5.30	65	35414	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.60%
32) Benzene-d6	5.33	84	115994	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.32	67	39564	5.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.40%
41) Toluene-d8	7.56	98	102130	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloropropene-	7.85	79	15253	4.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.60%
46) 2-Hexanone-d5	8.31	63	78856	50.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.80%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	34487	5.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.00%
64) 1,2-Dichlorobenzene-d4	11.85	152	42911	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds					Ovalue
13) Acetone	2.31	43	2400	1.707 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061419\
Data File : VU032740.D
Acq On : 14 Jun 2019 22:35
Operator : JC/SP
Sample : K3383-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0XK0

Quant Time: Jun 17 09:00:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 23:49:20 2019
Response via : Initial Calibration

