

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061419\
 Data File : VU032765.D
 Acq On : 15 Jun 2019 09:43
 Operator : JC/SP
 Sample : K3385-16
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COMB5

Quant Time: Jun 17 03:48:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 17 01:26:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	94843	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	90691	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	43623	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	24778	4.02	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.40%
7) Chloroethane-d5	1.67	69	23678	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.80%
11) 1,1-Dichloroethene-d2	2.27	63	40994	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	4.18	46	100470	58.38	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.76%
24) Chloroform-d	4.64	84	59577	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
26) 1,2-Dichloroethane-d4	5.30	65	33842	5.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.40%
32) Benzene-d6	5.33	84	110258	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.32	67	36628	5.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.40%
41) Toluene-d8	7.56	98	93882	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	7.85	79	13927	4.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.80%
46) 2-Hexanone-d5	8.31	63	74519	50.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.16%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	32737	5.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	39604	5.05	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.00%

Target Compounds					Ovalue
13) Acetone	2.32	43	4150	3.107 ug/L	89
31) Carbon tetrachloride	5.13	117	6835	0.680 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061419\
Data File : VU032765.D
Acq On : 15 Jun 2019 09:43
Operator : JC/SP
Sample : K3385-16
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COMB5

Quant Time: Jun 17 03:48:38 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 17 01:26:50 2019
Response via : Initial Calibration

