

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
 Data File : VU044183.D
 Acq On : 14 Jun 2021 15:56
 Operator : SY/MD
 Sample : M2672-27
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 17-B6(32-36)

Quant Time: Jun 15 02:46:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	158844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.260	114	293663	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.424	117	293515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	136417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	130776	51.771	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.540%	
35) Dibromofluoromethane	5.298	113	97309	47.635	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.280%	
50) Toluene-d8	7.906	98	386169	49.223	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.440%	
62) 4-Bromofluorobenzene	10.639	95	138399	44.020	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.040%	
Target Compounds						
					Qvalue	
16) Acetone	2.649	43	13220	8.063	ug/l	94
20) Methylene Chloride	3.057	84	9435	3.018	ug/l	93
43) Isopropyl Acetate	5.922	43	17955	3.235	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
Data File : VU044183.D
Acq On : 14 Jun 2021 15:56
Operator : SY/MD
Sample : M2672-27
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
17-B6(32-36)

Quant Time: Jun 15 02:46:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

