

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
 Data File : VU044187.D
 Acq On : 14 Jun 2021 17:28
 Operator : SY/MD
 Sample : M2672-31
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 17-B6(48-52)

Quant Time: Jun 15 02:47:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	159943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	296140	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	296265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	139289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	131896	51.856	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.720%	
35) Dibromofluoromethane	5.301	113	99205	48.157	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.320%	
50) Toluene-d8	7.906	98	392024	49.551	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.100%	
62) 4-Bromofluorobenzene	10.639	95	138402	43.653	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.300%	
Target Compounds						
						Qvalue
16) Acetone	2.649	43	13874	8.496	ug/l	99
20) Methylene Chloride	3.057	84	8084	2.385	ug/l	97
43) Isopropyl Acetate	5.922	43	18050	3.225	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
Data File : VU044187.D
Acq On : 14 Jun 2021 17:28
Operator : SY/MD
Sample : M2672-31
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
17-B6(48-52)

Quant Time: Jun 15 02:47:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

